

कार्यालय संयुक्त संचालक एवं अधीक्षक
गांधी स्मारक चिकित्सालय रीवा म.प्र.

निविदा



2024—2025

श्यामशाह चिकित्सा महाविद्यालय रीवा से सम्बद्ध सुपरस्पेशलिटी ब्लॉक के नेफ्रोलॉजी विभाग हेतु कंज्यूमेबल सामग्री के स्थानीय क्रय हेतु ई-निविदा

(लोकल पर्चेस)

(प्रथम आमंत्रण)

Handwritten signature and date: 27/10/24

निविदा क्रमांक-19/2024-25

कीमत रु. 2000/-
(रूपये दो हजार मात्र)

ONLINE TENDER SCHEDULE

ई -निविदा क्र- 19 / 2024-25

SEQ NO	SCHEDULE	START DATE & TIME	EXPIRY DATE & TIME
1	PURCHASE OF TENDER ONLINE	10.10.2024 11.00 PM	01.11.2024 06.00 PM
2	BID SUBMISSION ONLINE	10.10.2024 11.00 PM	01.11.2024 06.00 PM
3	TECHNICAL PROPOSAL OPEN	04.11.2024 11.00AM	-
4	FINANCIAL BID OPEN	पृथक से सूचना दी जावेगी।	-

ईमेल-: jdsupdtgmhrewa@rediffmail.com,

संपर्क-: 07662-350106, 9575285501, 9617348113, 9644695933, 7747078415

पोर्टल में ऑनलाईन अपलोड किए जाने वाले दस्तावेजों की सूची:-

1. धरोहर राशि रूपये 50,000.00 (रु. पचास हजार) ऑनलाईन जमा की रसीद।
2. मूल निविदा प्रपत्र हस्ताक्षर एवं सील सहित (Online purchased) सभी पृष्ठ
3. फर्म/संस्था की स्थापना का प्रमाण पत्र(गुमाश्ता)
4. 100 रु. के नॉन-ज्यूडिशियल स्टाम्प पर निर्धारित घोषणा पत्र
5. पैन कार्ड
6. वित्तीय वर्ष, 2022-23 एवं 2023-24 (एसेसमेंट ईयर 2023-24 एवं 2024-25) का आयकर रिटर्न जमा करने की स्वयं द्वारा सत्यापित छायाप्रति)
7. वार्षिक टर्न ओवर संबंधी प्रमाण पत्र सी.ए. द्वारा जारी। (रु 30.00 लाख वार्षिक से कम नहीं होना चाहिये वर्ष 2023-24)
8. जी.एस.टी. रजिस्ट्रेशन।
9. जीवित ड्रग लाइसेंस।
10. अधिकृत डीलर को निर्माता कंपनी द्वारा जारी अद्यतन अर्थराईजेशन पत्र की प्रति।
11. USFDA / European CE/DGCI सर्टिफाईड कंपनी का प्रमाण पत्र। यह प्रमाण पत्र सामग्री निर्धारित (Item Specific) उदाहरणतः सामग्री की सूची में दर्शाये गये विनिर्देशों के अनुरूप ही सर्टिफिकेशन मान्य होगा। अन्यथा अमान्य कर दिया जाएगा।
12. निविदा में प्रत्येक निविदाकार द्वारा भरे गए निविदा प्रपत्र में सभी सामग्रियों की सूची (जिनके लिए निविदा में भाग लिया गया है) संलग्न करना आवश्यक है, सूची के न होने की स्थिति में निविदा को अमान्य माना जाएगा।
13. सूची में सम्मिलित समस्त सामग्रियों के तकनीकी विश्लेषण (टेक्निकल स्पेशफिकेशन) हेतु समस्त सामग्रियों के कैंटलॉग संलग्न करना आवश्यक होगा।
14. सूची में सम्मिलित समस्त सामग्रियों के 01 नग सैम्पल उपलब्ध कराना आवश्यक होगा। निविदा में चयनित न होने की स्थिति में समस्त सैम्पल वापस कर दिये जाएंगे।

तकनीकी बिड हेतु प्रपत्र:-

स. क्र.	तकनीकी मापदण्ड हेतु अनिवार्य दस्तावेज	निविदाकार के द्वारा टिक किया जावे। पृष्ठ क. दर्शाये		परीक्षण समिति द्वारा टिक किया जावे।	
		हां	नहीं	हां	नहीं
1.	धरोहर राशि रूपये 50,000.00 (रु. पचास हजार) ऑनलाईन जमा की रसीद।				
2.	मूल निविदा प्रपत्र हस्ताक्षर एवं सील सहित (Online purchased) सभी पृष्ठ				
3.	फर्म/संस्था की स्थापना का प्रमाण पत्र(गुमाश्ता)				
4.	100 रु. के नॉन-ज्यूडिशियल स्टाम्प पर निर्धारित घोषणा पत्र				
5.	पैन कार्ड				
6.	वित्तीय वर्ष, 2022-23 एवं 2023-24 (एसेसमेंट ईयर 2023-24 एवं 2024-25) का आयकर रिटर्न जमा करने की स्वयं द्वारा सत्यापित छायाप्रति)				
7.	वार्षिक टर्न ओवर संबंधी प्रमाण पत्र सी.ए. द्वारा जारी। (रु 30.00 लाख वार्षिक से कम नहीं होना चाहिये वर्ष 2023-24)				
8.	जीवित ड्रग लाइसेंस।				
9.	जी.एस.टी. रजिस्ट्रेशन।				
10.	अधिकृत डीलर को निर्माता कंपनी द्वारा जारी अद्यतन अर्थराईजेशन पत्र की प्रति।				
11.	USFDA / European CE/DGCI सर्टिफाईड कंपनी का प्रमाण पत्र। यह प्रमाण पत्र सामग्री निर्धारित (Item Specific) उदाहरणतः सामग्री की सूची में दर्शाये गये विनिर्देशों के अनुरूप ही सर्टिफिकेशन मान्य होगा। अन्यथा अमान्य कर दिया जाएगा।				
12.	निविदा में प्रत्येक निविदाकार द्वारा भरे गए निविदा प्रपत्र में सभी सामग्रियों की सूची (जिनके लिए निविदा में भाग लिया गया है) संलग्न करना आवश्यक है, सूची के न होने की स्थिति में निविदा को अमान्य माना जाएगा।				
13.	सूची में सम्मिलित समस्त सामग्रियों के तकनीकी विश्लेषण (टेक्निकल स्पेशिफिकेशन) हेतु समस्त सामग्रियों के कैंटलॉग संलग्न करना आवश्यक होगा।				
14.	सूची में सम्मिलित समस्त सामग्रियों के 01 नमूने उपलब्ध कराना आवश्यक होगा। निविदा में चयनित न होने की स्थिति में समस्त नमूने वापस कर दिये जाएंगे।				

नोट:-तकनीकी मापदण्ड हेतु संलग्न किए गए दस्तावेजों के पृष्ठ क. दर्ज किया जाना अनिवार्य है।

निविदाकर्ता संस्था प्रमुख के हस्ताक्षर

निविदाकर्ता संस्था प्रमुख का नाम

(स्पष्ट अक्षरों)


संस्था का नाम एवं पूर्ण पता एवं सील


बैंक का नाम, खाता क. एंव आई.एफ.एस.कोड.....

श्यामशाह चिकित्सा महाविद्यालय रीवा से सम्बद्ध सुपरस्पेशलिटी ब्लॉक के नेफ्रोलॉजी विभाग हेतु कंज्यूमेबल सामग्री के स्थानीय क्रय (लोकल पर्चेस) (प्रथम आमंत्रण)

ई-निविदा फार्म नियम एवं शर्तें

1. श्यामशाह चिकित्सा महाविद्यालय रीवा से सम्बद्ध सुपरस्पेशलिटी ब्लॉक के नेफ्रोलॉजी विभाग हेतु कंज्यूमेबल सामग्री के स्थानीय क्रय (लोकल पर्चेस) (प्रथम आमंत्रण) आमंत्रित की जाती है। निविदा प्रपत्र वेबसाइट <https://mptenders.gov.in/> पर आनलाईन भुगतान कर क्रय किये जा सकेंगे। इसे चिकित्सा महाविद्यालय रीवा की वेबसाइट <http://www.ssmcrewa.com> में भी देखा जा सकेगा।
2. आनलाईन निविदा जमा करने की निर्धारित अवधि के पश्चात प्राप्त निविदाएँ स्वीकार नहीं की जावेगी। ऑनलाईन निविदा फार्म क्रय करने का मूल्य रु. 2000.00 निर्धारित है यह राशी वापस योग्य नहीं है।
3. निविदा के साथ अमानती राशि रूपये 50000.00 (रूपये पचास हजार मात्र) ऑनलाईन जमा करनी होगी। अमानत राशि पर किसी भी प्रकार का ब्याज देय नहीं होगा।
4. प्राइस बिड केवल ऑनलाईन ही भरी जावे तथा भारतीय मुद्रा में ही स्वीकार्य होगी।
5. कृपया प्राइस बिड भरने के पहले समस्त दस्तावेज एवं स्पेसिफिकेशन ध्यानपूर्वक पढ़ एवं समझ लें, उसके पश्चात ही निविदा भरें।
6. निविदाकार को ऑनलाईन क्रय किए हुए निविदा प्रपत्र एवं उसके साथ संलग्न कागजातों के प्रत्येक पृष्ठ पर अपने हस्ताक्षर कर सील लगाकर अपलोड करना होगा।
7. चिकित्सालय की भण्डार क्रय समिति को किसी भी अथवा समस्त निविदाओं को कारण बताकर अमान्य करने का अधिकार होगा।
8. सामान्यतः न्यूनतम मूल्य दर की निविदा को स्वीकार किया जावेगा, तथापि क्रय समिति शासनहित/जनहित में समग्र रूप से विचार करते हुये अन्यथा निर्णय लेने हेतु अधिकार सम्पन्न होगी।
9. निविदा एक वर्ष अथवा आगामी नवीन निविदा होने तक अथवा म.प्र. शासन के Procurement प्लान के अंतर्गत राज्य स्तरीय रेट कान्टेक्ट के प्रभावशील होने, जो भी पहले लागू हो जाये तक प्रभावी रहेगी।
10. निविदा केवल निर्माता कंपनी अथवा उनके अधिकृत डीलर द्वारा ही भरी जाए।
11. सामग्री विलम्ब या आधी अधूरी प्रदाय करने की स्थिति में द्वितीय निविदाकार (एल-02) से क्रय कर जो राशी का अंतर होगा वह प्रथम निविदाकार की प्रतिभूति या देयक से काटा जावेगा।
12. प्रदाय करते हुये इस बात का ध्यान रखा जावे की कंज्यूमेबल/सर्जिकल सामग्री उच्च क्वालिटी (DGCI/ European CE /USFDA) जो कि टेन्डर प्रपत्र में विवरण एवं आवश्यक सुरक्षा जानकारी के साथ अंकित है की ही दी जावें। कोविड महामारी को देखते हुए सामग्री क्रय आदेश जारी करने के दिनांक से 07 दिवस के अंदर चिकित्सालय के भण्डार में पहुंचाना अनिवार्य होगी।



SUPERINTENDENT
Super Speciality Hospital
Shyam Shah Medical College, Rewa


Dr. Rohan Dwivedi
Associate Professor
HOD, Nephrology
SSB & SSMC Rewa (M.P.)
Reg. No. MP16020

13. अनुबंध करने के बाद कार्य प्रारम्भ नहीं करने की स्थिति में जमा प्रतिभूति राशि राजसात कर आदेश निरस्त करने का अधिकार अधिष्ठाता/संयुक्त संचालक एवं अधीक्षक को रहेगा तथा कार्यादेश अन्य संस्था, द्वितीय निविदाकार (एल-02) को दिया जा सकेगा इस कारण होने वाली वित्तीय हानि को भी राजकीय हानि मानते हुये निरस्त निविदाकर्ता से वसूली की जावेगी।
14. आकस्मिक आवश्यकता पड़ने पर अधीक्षक/प्राधिकृत अधिकारी द्वारा दूरभाष पर दिए गए निर्देशों पर प्रदायगी करना अनिवार्य होगा।
15. किसी भी विवाद, व्यवधान अथवा लागू नियम व शर्तों को समझने में दुविधा की स्थिति में मित्रवत शालीन व्यवहार करते हुये संयुक्त संचालक एवं अधीक्षक गांधी स्मारक चिकित्सालय से उनके कार्यालय में चर्चा कर समाधान निकालने का प्रयास करना होगा चर्चा उपरांत पहुंचे समाधान को निविदाकार को मान्य करना होगा।
16. निविदा के साथ दुकान का वैध ड्रग लाइसेंस तथा आयकर रिटर्न वित्तीय वर्ष 2022-23, एवं 2023-24 की छायाप्रति लगानी होगी। वार्षिक टर्न ओवर वर्ष 2023-24 तीस लाख (30 LAKH) से कम नहीं होना चाहिये इसके लिये सी.ए. द्वारा जारी प्रमाण पत्र संलग्न करना अनिवार्य होगा।
17. मुख्य कार्यपालन अधिकारी एवं अधिष्ठाता चिकि.महा. रीवा के पास किसी भी निविदा को ग्राह्य करने अथवा नहीं करने तथा साथ ही जनहित में निविदा शर्तों में परिवर्तन करने का अधिकार सुरक्षित रहेगा।
18. निविदाकार को निविदा प्रपत्र की समस्त शर्तें मंजूर करनी होगी, जिन्हें पालन करने के लिये वह वैधानिक रूप से बाध्य है।
19. इस निविदा में स्वीकृत दरों से केवल उन्ही कंज्यूमेबल/सर्जिकल सामग्रियों का क्रय किया जावेगा जिनकी दरें एम.पी.एच.एच.सी.एल की सूची में उपलब्ध नहीं हैं अथवा उनके द्वारा समय से प्रदायगी नहीं की जा रही है अथवा स्थानीय स्तर पर अनुबंधित फर्मों द्वारा प्रदायगी नहीं की जा रही है और मरीजों के उपचार हेतु तात्कालिक रूप से आवश्यक है।
20. प्रदायकर्ता फर्म को प्रदाय की गई कंज्यूमेबल/सर्जिकल सामग्री की एन.ए.बी.एल. /इनहाउस रिपोर्ट प्रदायगी के साथ उपलब्ध करानी होगी।
21. जिस निविदाकार की दरें स्वीकृत होगी उन्हें सामग्री प्रदाय करना अनिवार्य होगा।
22. निविदा के सम्बन्ध में न्यायिक विवाद रीवा स्थित न्यायालय के क्षेत्राधिकार के अंतर्गत होगा।

संयुक्त संचालक एवं अधीक्षक
गांधी स्मारक चिकित्सालय रीवा (म.प्र.)


Dr. Rohan Dwivedi
Associate Professor
HOD, Nephrology
SSB & SSMC Rewa (M.P.)
Reg. No. MP16020


SUPERINTENDENT
Super Speciality Hospital
Shyam Shah Medical College, Rewa

घोषणा पत्र (शपथ पत्र)

(100/- रुपये के नान-ज्यूडिसियल स्टाम्प पेपर पर प्रस्तुत किया जाये)

1. मैं/हम भारत का/के नागरिक हूँ/हैं।
2. मैं/हम भारतीय संविदा अधिनियम के अनुसार निविदा कार्य के लिए सक्षम हूँ/हैं।
3. मुझे/हमें म.प्र. शासन के किसी भी विभाग द्वारा निविदाकार्य से बहिष्कृत नहीं किया गया है।
4. मैं/हम किसी ऐसे कान्ट्रैक्टर, जो कि म.प्र. शासन के किसी विभाग द्वारा बहिष्कृत किया गया हो, न तो अभिकर्ता हूँ/हैं और न ही उसके लिये कार्य करता हूँ/करते हैं।
5. मेरे/हमारे द्वारा निविदा की शर्तों के अलावा अन्य कोई शर्तें प्रस्तुत नहीं की गयी है। यदि अन्य कोई शर्तें प्रस्तुत की गयी है, तो उसे अमान्य माना जाये।
6. श्यामशाह चिकित्सा महाविद्यालय/संजय गांधी स्मृति चिकित्सालय/गांधी स्मारक चिकित्सालय रीवा के किसी अधिकारी/कर्मचारी से संबंध नहीं है।
7. हमारी संस्था के विरुद्ध सी.बी.आई./लोकायुक्त/ई.ओ.डब्ल्यू/शासकीय कार्यालय इत्यादि में कोई जाचं लबित नहीं है और न ही सक्षम न्यायालय में आपराधिक प्रकरण दर्ज है और न ही दण्डित किया गया है।
8. केंद्र/राज्य सरकार कोई भी कर मेरी संस्था/मेरे उपर बकाया नहीं है।
9. मेरे द्वारा प्रस्तुत जानकारी सही है। असत्य पाये जाने पर मेरे विरुद्ध विधिसम्मत कार्यवाही के किए जाने पर मैं स्वयं उत्तरदायी रहूंगा।

गवाह के हस्ताक्षर

निविदाकार के हस्ताक्षर


1. गवाह का नाम.....
पता.....

संस्था के समक्ष अधिकारी का नाम.....
पता.....

2. गवाह का नाम.....

संस्था का नाम एवं पता.....

पता


Dr. Rohan Dwivedi
Associate Professor
HOD, Nephrology
SSB & SSMC Rewa (M.P.)
Reg. No. MP16020


SUPERINTENDENT
Super Speciality Hospital
Shyam Shah Medical College, Rewa

Rates to be filled online.

S.No.	Description of consumable	Make(Mandatory)	Rate	GST	Total
1	Blood tubing set compatible with Fresenius 5008s dialysis machine Adult size USFDA/ EUROPEAN CE approved and DCGI marked				
2	Blood tubing set compatible with Fresenius 5008s dialysis machine pediatric size				
3	Filter compatible with 5008s dialysis machine (Diasafe)USFDA/ EUROPEAN CE approved and DCGI marked				
4	Filter compatible with NiproSurdial 55 plus (Ultra filter)USFDA/ EUROPEAN CE approved and DCGI marked				
5	Low flux dialyzers SA-0.4 m2; Polysulfone membrane, polycarbonate-housing, Sterilization Steam Uf co1.7, urea cl 183 @ Qb-300 USFDA/ EUROPEAN CE approved and DCGI marked				
6	Low flux dialyzers SA-0.6 m2; Polysulfone membrane, polycarbonate-housing, Sterilization Steam ufco8, urea 210Qb-300 USFDA/ EUROPEAN CE approved and DCGI marked				
7	Plasma filter, Hollow fiber, Synthetic membrane, 0.6m2 USFDA/ EUROPEAN CE approved and DCGI marked				
8	Plasma filter, Hollow fiber, Synthetic membrane, 0.3m2 USFDA/ EUROPEAN CE approved and DCGI marked				
9	Disposable Dressing Kit for dialysis initiation - comprising of presterilized tray,1-2 cups , non-woven Absorbent sheet, conversing sheet, ETO sterile Indicator pouch each 1 No, Gauze swabs - 6-8 Nos, Cotton balls-5-7 Nos				
10	Adhesive dressing for dialysis application				
11	Adhesive plaster dressing for dialysis application				
12	Adhesive transparent dressing for Dialysis Application 8.5-10 cm				
13	Chitosan Hemostatic Dressing (size2x2cm)				
14	Sterile Transparent Ultrasound probe cover				
15	Double Lumen Dialysis Curved extension PEDIATRIC catheter Kit with following parameters 1. Two lumens of Double -D design. 2. Laser Cut Side slots to avoid positional occlusion. 3. Clear Silicon Extensions. 4. Size 8.5/9 Fr x 12 cm.5. 2 dilators in a pack 6. introducer needle 16G 7.spring guide wire J type, 8. suture with curved needle, 9. scalpel blade 10. end caps USFDA DCGI marked				


SUPERINTENDENT
Super Speciality Hospital
Shyam Shah Medical College, Rewa


8

Dr. Rohan Dwivedi
 Associate Professor
 HOD, Nephrology
 SSB & SSMC Rewa (M.P.)
 Reg. No. MP16020

16	Double Lumen Dialysis curved extension catheter Kit with following parameters 1. Two lumens of Double-D design. 2. Laser Cut Side slots to avoid positional occlusion. 3. Clear Silicon Extensions. 4. Size 12 Fr x 20 cm 5. 2 dilators in a pack 6. introducer needle 18G 7.spring guide wire J type, 8.suturewith curved needle, 9.scalpel blade USFDA DCGI marked				
17	Triple Lumen Dialysis Curved extension catheter Kit with following parameters 1. Three lumens of Modified Double-D design 2. Laser Cut Side slots to avoid positional occlusion 3. Clear Silicon Extensions 4. sizes 12 Fr x 13.5 cm/12 Fr x 16 cm 5. Should have- 2 dilators in the pack6. introducer needle 18G 7.spring guide wire J type, 8.suturewith curved needle, 9.scalpel blade				
18	Double lumen symmetrical tip tunneled hemodialysis catheter kit 1. Two lumens of Double-D design with symmetrical tip 2. Spiral Z tip design with Laser Cut Side Slots 3. Carbothane material 4. sizes 14.5 - 16 Fr with different lengths 19cm, 23 cm, 28cm, 55cm from tip to cuff 5. Vessel dilator with Flow guardto avoid blood spillage or air embolism 6. introducer needle 18G 7. spring guide wire J type, 8. suturewith curved needle, 9. scalpel blade 10. end cap 11. adhesive dressings 2 each USFDA DCGI marked				
19	Double Lumen split tip tunneled hemodialysis catheter kit Adult size 1. Catheter size 14,5Fr- 16 Fr 2. length 19, 23, 27,31,35 from tip to cuff 3. AIRGUARD Valved Introducer with Peel-Away Sheath 4. Dilator Dilator – 8F Dilator – 10-12 F 5. J-Tip Guidewire – 0.038 in. 6. Introducer Needle – 18 Gauge 7. Tunneler 8. Adhesive Dressings-2 each 9. End Cap 10. suturewith curved needle, 11. scalpel blade USFDA DCGI marked				
20	Double Lumen split tip tunneled hemodialysis catheter kit pediatric size 1. Tip to cuff length 8, 10, 12, 15, 19, 23 cm 2. 7.5Fr-10 Fr Catheter 3. 8F-10Fr AirGuard Valved Introducer with Peel Away Sheath/Dilator 4. Dilator 5. Tunneler 6. 2 End Caps 7. J-Tip Guidewire 0.032" 8. 18 Gauge Introducer Needle 9. 2 Adhesive Dressing USFDA DCGI marked				


SUPERINTENDENT
Super Speciality Hospital
Shyam Shah Medical College, Rewa


Dr. Rohan Dwivedi
Associate Professor
HOD, Nephrology
SSB & SSMC, Rewa (M.P.)
Reg. No. MP16020



21	Disposable Core tissue Biopsy needle compatible with Magnum Biopsy instrument size: 16G X 16Cm USFDA DCGI marked				
22	Disposable Core tissue Biopsy needle compatible with Magnum Biopsy instrument size: 18G X 16Cm USFDA DCGI marked				
23	Disposable Core tissue Biopsy needle compatible with Magnum Biopsy instrument size: 18G X 10Cm USFDA/ EUROPEAN CE approved and DCGI marked				
24	Sterilit oil spray to reduce friction and corrosion due to friction for instruments (300ml)				
25	Bacterial Filter for Macquet Ventilator				
26	RBC Leukocyte reduction filters (lab side with transfer bag) 3rd generation				
27	Micro-puncture kit				
28	PTA & INJECTION COMBINED · 1 Should have 3 port system. · 2 Should be available in sizes, Dia(5mm,6mm,7mm) with nominal pressure of 14atm upto burst pressure of 25atm. · 3 Should enable simultaneous angioplasty and variety of injections at any time. · 4 Should combine the functionality of high-pressure balloon catheter and diagnostic catheter · 5 Should be able to perform an entire procedure both angioplasty and injection of diagnostic or therapeutic fluids. 6 Should be able to perform reflex angiography. 7 The balloon should have supervision Technology USFDA/ EUROPEAN CE approved and DCGI marked				
29	Tenckhoff peritoneal dialysis catheter, 2cuffs , 42cm, Straight with introducer set for percutaneous insertion USFDA/ EUROPEAN CE approved and DCGI marked				
30	Tenckhoff peritoneal dialysis catheter, 2cuffs , 31cm, Straight with introducer set for percutaneous insertion USFDA/ EUROPEAN CE approved and DCGI marked				
31	Tenckhoff peritoneal dialysis catheter, 2cuffs , 37cm, Straight with introducer set for percutaneous insertion USFDA/ EUROPEAN CE approved and DCGI marked				
32	Tenckhoff peritoneal dialysis catheter, 2cuffs , 46cm, Straight with introducer set for percutaneous insertion USFDA/ EUROPEAN CE approved and DCGI marked				

SUPERINTENDENT
Super Speciality Hospital
Shyam Shah Medical College, Rewa

Dr. Rohan Dwivedi
 Associate Professor
 HOD, Nephrology
 SSB & SSMC Rewa (M.P.)
 Reg. No. MP16020

33	Tenckhoff peritoneal dialysis catheter, single cuffs , 31cm, Straight with introducer set for percutaneous insertion USFDA/ EUROPEAN CE approved and DCGI marked				
34	Tenckhoff peritoneal dialysis catheter, single cuffs , 37cm, Straight with introducer set for percutaneous insertion USFDA/ EUROPEAN CE approved and DCGI marked				
35	Tenckhoff peritoneal dialysis catheter, single cuffs , 42cm, Straight with introducer set for percutaneous insertion USFDA/ EUROPEAN CE approved and DCGI marked				
36	Acute semirigid peritoneal dialysis catheter with Y transfer set Adult size USFDA/ EUROPEAN CE approved and DCGI marked				
37	Acute semirigid peritoneal dialysis catheter with Y transfer set pediatric size USFDA/ EUROPEAN CE approved and DCGI marked				
38	Pediatric acute peritoneal dialysis set Including 1. Polyurethane catheter catheter- 8.5 fr, length 8 cm 2. Guide wire -0.35 inch length -30 cm 3. introducer needle 18T gage 4. connecting tube-30 cm 5. dilator 6. retention disc with the pull silicone USFDA/ EUROPEAN CE approved and DCGI marked				
39	Locking titanium adapter for peritoneal dialysis catheter USFDA/ EUROPEAN CE approved and DCGI marked				
40	Minicap extended life PD transfer set with twist clamp USFDA/ EUROPEAN CE approved and DCGI marked				
41	Peritoneal dialysis fluids 1.5 % 2000ml bag Each 100 ml – dextrose anhydrous 1.5gm, sodium lactate 448 mg, Na Cl 538 mg, Cacl 25.7 mg, Mgcl- 5.08 mg with integrated asymmetric Y set with or without minicap with colour code				
42	Peritoneal dialysis fluids 2.5 % 2000ml bag Each 100 ml – dextrose anhydrous 2.5gm, sodium lactate 448 mg, Na Cl 538 mg, Cacl 25.7 mg, Mgcl- 5.08 mg with integrated asymmetric Y set with or without minicap with colour code				
43	Peritoneal dialysis fluids 7.5 % 2000ml bag Each 100 ml – icodextrin 7.5gm, sodium lactate 448 mg, Na Cl 538 mg, Cacl 25.7 mg, Mgcl- 5.08 mg with water for injection Q S				
44	Peritoneal dialysis fluids 1.5 % 5000ml bag Each 100 ml – dextrose anhydrous 1.5gm, sodium lactate 448 mg, Na Cl 538 mg, Cacl 25.7 mg, Mgcl- 5.08 mg with				


45	Peritoneal dialysis fluids 2.5 % 5000ml bag Each 100 ml – dextrose anhydrous 2.5gm, sodium lactate 448 mg, Na Cl 538 mg, Cacl 25.7 mg, Mgcl- 5.08 mg with water for injection Q S				
46	16 Fr Tearway introducer sheath for Peritoneal Dialysis				
47	Multifiltrate cassette compatible with Fresenius CRRT machine USFDA/ EUROPEAN CE approved and DCGI marked				
48	Substitution line compatible with Fresenius CRRT machine USFDA/ EUROPEAN CE approved and DCGI marked				
49	Dialysate line compatible with Fresenius CRRT machine USFDA/ EUROPEAN CE approved and DCGI marked				
50	Ultra-flux filter compatible with Fresenius CRRT machine USFDA/ EUROPEAN CE approved and DCGI marked				
51	Filtrate bag compatible with Fresenius CRRT machine USFDA/ EUROPEAN CE approved and DCGI marked				
52	Female spike adaptor 4 each compatible with Fresenius CRRT machine USFDA/ EUROPEAN CE approved and DCGI marked				
53	Solution for continuous hemofiltration and hemodialysis, potassium free, Bicarbonate 30-35 meq/L, Calcium 2.5 - 3.5 meq/L with Dextrose, 5000ml/bag USFDA/ EUROPEAN CE approved and DCGI marked				
54	Solution for continuous hemofiltration and hemodialysis, potassium free, Bicarbonate 30-35 meq/L, Calcium 2.5 - 3.5 meq/L without Dextrose, 5000ml/bag USFDA/ EUROPEAN CE approved and DCGI marked				
55	Solution for continuous hemofiltration and hemodialysis, potassium free, Bicarbonate 30-35 meq/L, with Dextrose, without calcium, 5000ml/bag USFDA/ EUROPEAN CE approved and DCGI marked				

गवाह के हस्ताक्षर

1. गवाह का नाम.....
पता.....

2. गवाह का नाम.....

पता


SUPERINTENDENT
Super Speciality Hospital
Shyam Shah Medical College, Rewa

निविदाकार के हस्ताक्षर

संस्था के समक्ष अधिकारी का नाम.....
पता.....

संस्था का नाम एवं पता.....

Dr. Rohan Dwivedi
Associate Professor
HOD Nephrology
SSB & SSMC Rewa (M.P.)
Reg. No. MP-16020


Technical Bid Mandatory Document to be filled and attached along with tender document.

No.	Description of consumable	Pack Size	Unit	Pack size offered	Name of Brand/Company Quoted	Indian/ Imported/ Country of Origin	Catalogues Yes/No	Sample Yes/No
1	Blood tubing set compatible with Fresenius 5008s dialysis machine Adult size USFDA/ EUROPEAN CE approved and DCGI marked							
2	Blood tubing set compatible with Fresenius 5008s dialysis machine pediatric size							
3	Filter compatible with 5008s dialysis machine (Diasafe)USFDA/ EUROPEAN CE approved and DCGI marked							
4	Filter compatible with NiproSurdial 55 plus (Ultra filter)USFDA/ EUROPEAN CE approved and DCGI marked							
5	Low flux dialyzers SA-0.4 m2; Polysulfone membrane, polycarbonate-housing, Sterilization Steam Uf co1.7, urea cl 183 @ Qb-300 USFDA/ EUROPEAN CE approved and DCGI marked							
6	Low flux dialyzers SA-0.6 m2; Polysulfone membrane, polycarbonate-housing, Sterilization Steam ufco8, urea 210Qb-300 USFDA/ EUROPEAN CE approved and DCGI marked							
7	Plasma filter, Hollow fiber, Synthetic membrane, 0.6m2 USFDA/ EUROPEAN CE approved and DCGI marked							
8	Plasma filter, Hollow fiber, Synthetic membrane, 0.3m2 USFDA/ EUROPEAN CE approved and DCGI marked							
9	Disposable Dressing Kit for dialysis initiation - comprising of presterilized tray,1-2 cups , non-woven Absorbent sheet, conversing sheet, ETO sterile Indicator pouch each 1 No, Gauze swabs -6-8 Nos, Cotton balls-5-7 Nos							
10	Adhesive dressing for dialysis application							
11	Adhesive plaster dressing for dialysis application							
12	Adhesive transparent dressing for Dialysis Application 8.5-10 cm							
13	Chitosan Hemostatic Dressing (size2x2cm)							
14	Sterile Transparent Ultrasound probe cover							
15	Double Lumen Dialysis Curved extension PEDIATRIC catheter Kit with following parameters 1. Two lumens of Double -D design. 2. Laser Cut Side slots to avoid positional occlusion. 3. Clear Silicon Extensions. 4. Size 8.5/9 Fr x 12 cm.5. 2 dilators in a pack 6. introducer needle 16G 7.spring guide wire J type, 8.suturewith curved needle, 9.scalpel blade 10. end caps USFDA DCGI marked							
16	Double Lumen Dialysis curved extension catheter Kit with following parameters 1. Two lumens of Double-D design. 2. Laser Cut Side slots to avoid positional occlusion. 3. Clear Silicon Extensions. 4. Size 12 Fr x 20 cm 5. 2 dilators in a pack 6. introducer needle 18G 7.spring guide wire J type, 8.suturewith curved needle, 9 scalpel blade USFDA DCGI marked							
17	Triple Lumen Dialysis Curved extension catheter Kit with following parameters 1. Three lumens of Modified Double-D design 2. Laser Cut Side slots to avoid positional occlusion 3. Clear Silicon Extensions 4. sizes 12 Fr x 13.5 cm/12 Fr x 16 cm 5. Should have- 2 dilators in the pack6. introducer needle 18G 7.spring guide wire J type, 8.suturewith curved needle, 9.scalpel blade							
18	Double lumen symmetrical tip tunneled hemodialysis catheter kit 1. Two lumens of Double-D design with symmetrical tip 2. Spiral Z tip design with Laser Cut Side Slots 3. Carbothane material 4. sizes 14.5 -16 Fr with different lengths 19cm, 23 cm, 28cm, 55cm from tip to cuff 5. Vessel dilator with Flow guardto avoid blood spillage or air embolism 6. introducer needle 18G 7. spring guide wire J type, 8. suturewith curved needle, 9. scalpel blade 10. end cap 11. adhesive dressings 2 each USFDA DCGI marked							
19	Double Lumen split tip tunneled hemodialysis catheter kit Adult size 1. Catheter size 14,5Fr- 16 Fr 2. length 19, 23, 27,31,35 from tip to cuff 3. AIRGUARD Valved Introducer with Peel-Away Sheath 4. Dilator Dilator – 8F Dilator – 10-12 F 5. J-Tip Guidewire – 0.038 in. 6. Introducer Needle – 18 Gauge 7. Tunneler 8. Adhesive Dressings-2 each 9. End Cap 10. suturewith curved needle, 11. scalpel blade USFDA DCGI marked							
20	Double Lumen split tip tunneled hemodialysis catheter kit pediatric size 1. Tip to cuff length 8, 10, 12, 15, 19, 23 cm 2. 7.5Fr-10 Fr Catheter 3. 8F-10Fr AirGuard Valved Introducer with Peel Away Sheath/Dilator 4. Dilator 5. Tunneler 6. 2 End Caps 7. J-Tip Guidewire 0.032" 8. 18 Gauge Introducer Needle 9. 2 Adhesive Dressing USFDA DCGI marked							
21	Disposable Core tissue Biopsy needle compatible with Magnum Biopsy instrument size: 16G X 16Cm USFDA DCGI marked							
22	Disposable Core tissue Biopsy needle compatible with Magnum Biopsy instrument size: 18G X 16Cm USFDA DCGI marked							
23	Disposable Core tissue Biopsy needle compatible with Magnum Biopsy instrument size: 18G X 10Cm USFDA/ EUROPEAN CE approved and DCGI marked							
24	Sterilit oil spray to reduce friction and corrosion due to friction for instruments (300ml)							
25	Bacterial Filter for Macquet Ventilator							
26	RBC Leukocyte reduction filters (lab side with transfer bag) 3rd generation							
27	Micro-puncture kit							
28	PTA & INJECTION COMBINED · 1 Should have 3 port system. · 2 Should be available in sizes, Dia(5mm,6mm,7mm) with nominal pressure of 14atm upto burst pressure of 25atm. · 3 Should enable simultaneous angioplasty and variety of injections at any time. · 4 Should combine the functionality of high-pressure balloon catheter and diagnostic catheter · 5 Should be able to perform an entire procedure both angioplasty and injection of diagnostic or therapeutic fluids. 6 Should, be able to perform reflex angiography. 7 The balloon should have supervision Technology USFDA/ EUROPEAN CE approved and DCGI marked							

Dr. Rohan Dwivedi
Associate Professor
HOD, Nephrology
SSB & SSMC Rewa (M.P.)
Reg.No. MP16020

SUPERINTENDENT
Super Speciality Hospital
Shyam Shah Medical College, Rewa

29	Tenckhoff peritoneal dialysis catheter, 2cuffs, 42cm, Straight with introducer set for percutaneous insertion USFDA/ EUROPEAN CE approved and DCGI marked								
30	Tenckhoff peritoneal dialysis catheter, 2cuffs, 31cm, Straight with introducer set for percutaneous insertion USFDA/ EUROPEAN CE approved and DCGI marked								
31	Tenckhoff peritoneal dialysis catheter, 2cuffs, 37cm, Straight with introducer set for percutaneous insertion USFDA/ EUROPEAN CE approved and DCGI marked								
32	Tenckhoff peritoneal dialysis catheter, 2cuffs, 46cm, Straight with introducer set for percutaneous insertion USFDA/ EUROPEAN CE approved and DCGI marked								
33	Tenckhoff peritoneal dialysis catheter, single cuffs, 31cm, Straight with introducer set for percutaneous insertion USFDA/ EUROPEAN CE approved and DCGI marked								
34	Tenckhoff peritoneal dialysis catheter, single cuffs, 37cm, Straight with introducer set for percutaneous insertion USFDA/ EUROPEAN CE approved and DCGI marked								
35	Tenckhoff peritoneal dialysis catheter, single cuffs, 42cm, Straight with introducer set for percutaneous insertion USFDA/ EUROPEAN CE approved and DCGI marked								
36	Acute semirigid peritoneal dialysis catheter with Y transfer set Adult size USFDA/ EUROPEAN CE approved and DCGI marked								
37	Acute semirigid peritoneal dialysis catheter with Y transfer set pediatric size USFDA/ EUROPEAN CE approved and DCGI marked								
38	Pediatric acute peritoneal dialysis set Including 1. Polyurethane catheter 8.5 fr, length 8 cm 2. Guide wire -0.35 inch length -30 cm 3. introducer needle 18T gage 4. connecting tube-30 cm 5. dilator 6. retention disc with the pull silicone USFDA/ EUROPEAN CE approved and DCGI marked								
39	Locking titanium adapter for peritoneal dialysis catheter USFDA/ EUROPEAN CE approved and DCGI marked								
40	Minicap extended life PD transfer set with twist clamp USFDA/ EUROPEAN CE approved and DCGI marked								
41	Peritoneal dialysis fluids 1.5 % 2000ml bag Each 100 ml – dextrose anhydrous 1.5gm, sodium lactate 448 mg, Na Cl 538 mg, Cacl 25.7 mg, Mgcl- 5.08 mg with integrated asymmetric Y set with or without minicap with colour code								
42	Peritoneal dialysis fluids 2.5 % 2000ml bag Each 100 ml – dextrose anhydrous 2.5gm, sodium lactate 448 mg, Na Cl 538 mg, Cacl 25.7 mg, Mgcl- 5.08 mg with integrated asymmetric Y set with or without minicap with colour code								
43	Peritoneal dialysis fluids 7.5 % 2000ml bag Each 100 ml – icodextrin 7.5gm, sodium lactate 448 mg, Na Cl 538 mg, Cacl 25.7 mg, Mgcl- 5.08 mg with water for injection Q.S								
44	Peritoneal dialysis fluids 1.5 % 5000ml bag Each 100 ml – dextrose anhydrous 1.5gm, sodium lactate 448 mg, Na Cl 538 mg, Cacl 25.7 mg, Mgcl- 5.08 mg with								
45	Peritoneal dialysis fluids 2.5 % 5000ml bag Each 100 ml – dextrose anhydrous 2.5gm, sodium lactate 448 mg, Na Cl 538 mg, Cacl 25.7 mg, Mgcl- 5.08 mg with water for injection Q.S								
46	16 Fr Tearway introducer sheath for Peritoneal Dialysis								
47	Multifiltrate cassette compatible with Fresenius CRRT machine USFDA/ EUROPEAN CE approved and DCGI marked								
48	Substitution line compatible with Fresenius CRRT machine USFDA/ EUROPEAN CE approved and DCGI marked								
49	Dialysate line compatible with Fresenius CRRT machine USFDA/ EUROPEAN CE approved and DCGI marked								
50	Ultra-flux filter compatible with Fresenius CRRT machine USFDA/ EUROPEAN CE approved and DCGI marked								
51	Filtrate bag compatible with Fresenius CRRT machine USFDA/ EUROPEAN CE approved and DCGI marked								
52	Female spike adaptor 4 each compatible with Fresenius CRRT machine USFDA/ EUROPEAN CE approved and DCGI marked								
53	Solution for continuous hemofiltration and hemodialysis, potassium free, Bicarbonate 30-35 meq/L, Calcium 2.5 - 3.5 meq/L with Dextrose, 5000ml/bag USFDA/ EUROPEAN CE approved and DCGI marked								
54	Solution for continuous hemofiltration and hemodialysis, potassium free, Bicarbonate 30-35 meq/L, Calcium 2.5 - 3.5 meq/L without Dextrose, 5000ml/bag USFDA/ EUROPEAN CE approved and DCGI marked								
55	Solution for continuous hemofiltration and hemodialysis, potassium free, Bicarbonate 30-35 meq/L, with Dextrose, without calcium, 5000ml/bag USFDA/ EUROPEAN CE approved and DCGI marked								


Dr. Rohan Dwivedi
 Associate Professor
 HOD, Nephrology
 SSB & SSMC Rewa (M.P.)
 Reg. No. MP16020


SUPERINTENDENT
 Super Speciality Hospital
 Shyam Shah Medical College, Rewa

Technical Bid Mandatory Document to be filled and attached along with tender document.

S.No	Description of consumable	Pack Size	Unit	Pack size offered	Name of Brand/Company Quoted	Indian/ Imported/ Country of Origin	Catalogues Yes/No	Sample Yes/No
1	Blood tubing set compatible with Fresenius 5008s dialysis machine Adult size USFDA/ EUROPEAN CE approved and DCGI marked							
2	Blood tubing set compatible with Fresenius 5008s dialysis machine pediatric size							
3	Filter compatible with 5008s dialysis machine (Diasafe)USFDA/ EUROPEAN CE approved and DCGI marked							
4	Filter compatible with NiproSurdial 55 plus (Ultra filter)USFDA/ EUROPEAN CE approved and DCGI marked							
5	Low flux dialyzers SA-0.4 m2; Polysulfone membrane, polycarbonate-housing, Sterilization Steam Uf co1.7, urea cl 183 @ Qb-300 USFDA/ EUROPEAN CE approved and DCGI marked							
6	Low flux dialyzers SA-0.6 m2; Polysulfone membrane, polycarbonate-housing, Sterilization Steam ufco8, urea 210Qb-300 USFDA/ EUROPEAN CE approved and DCGI marked							
7	Plasma filter, Hollow fiber, Synthetic membrane, 0.6m2 USFDA/ EUROPEAN CE approved and DCGI marked							
8	Plasma filter, Hollow fiber, Synthetic membrane, 0.3m2 USFDA/ EUROPEAN CE approved and DCGI marked							
9	Disposable Dressing Kit for dialysis initiation - comprising of presterilized tray,1-2 cups , non-woven Absorbent sheet, conversing sheet, ETO sterile Indicator pouch each 1 No, Gauze swabs -6-8 Nos, Cotton balls-5-7 Nos							
10	Adhesive dressing for dialysis application							
11	Adhesive plaster dressing for dialysis application							
12	Adhesive transparent dressing for Dialysis Application 8.5-10 cm							
13	Chitosan Hemostatic Dressing (size2x2cm)							
14	Sterile Transparent Ultrasound probe cover							
15	Double Lumen Dialysis Curved extension PEDIATRIC catheter Kit with following parameters 1. Two lumens of Double -D design. 2. Laser Cut Side slots to avoid positional occlusion. 3. Clear Silicon Extensions. 4. Size 8.5/9 Fr x 12 cm.5. 2 dilators in a pack 6. introducer needle 16G 7.spring guide wire J type, 8.suturewith curved needle, 9.scalpel blade 10. end caps USFDA DCGI marked							
16	Double Lumen Dialysis curved extension catheter Kit with following parameters 1. Two lumens of Double-D design. 2. Laser Cut Side slots to avoid positional occlusion. 3. Clear Silicon Extensions. 4. Size 12 Fr x 20 cm 5. 2 dilators in a pack 6. introducer needle 18G 7.spring guide wire J type, 8.suturewith curved needle, 9.scalpel blade USFDA DCGI marked							
17	Triple Lumen Dialysis Curved extension catheter Kit with following parameters 1. Three lumens of Modified Double-D design 2. Laser Cut Side slots to avoid positional occlusion 3. Clear Silicon Extensions 4. sizes 12 Fr x 13.5 cm/12 Fr x 16 cm 5. Should have- 2 dilators in the pack6. introducer needle 18G 7.spring guide wire J type, 8.suturewith curved needle, 9.scalpel blade							
18	Double lumen symmetrical tip tunneled hemodialysis catheter kit 1. Two lumens of Double-D design with symmetrical tip 2. Spiral Z tip design with Laser Cut Side Slots 3. Carbothane material 4. sizes 14.5 -16 Fr with different lengths 19cm, 23 cm, 28cm, 55cm from tip to cuff 5. Vessel dilator with Flow guardto avoid blood spillage or air embolism 6. introducer needle 18G 7. spring guide wire J type, 8. suturewith curved needle, 9. scalpel blade 10. end cap 11. adhesive dressings 2 each USFDA DCGI marked							

S.No	Description of consumable	Pack Size	Unit	Pack size offered	Name of Brand/Company Quoted	Indian/Imported/Country of Origin	Catalogues Yes/No	Sample Yes/No
19	Double Lumen split tip tunneled hemodialysis catheter kit Adult size 1. Catheter size 14,5Fr- 16 Fr 2. length 19, 23, 27,31,35 from tip to cuff 3. AIRGUARD Valved Introducer with Peel-Away Sheath 4. Dilator Dilator – 8F Dilator – 10-12 F 5. J-Tip Guidewire – 0.038 in. 6. Introducer Needle – 18 Gauge 7. Tunneler 8. Adhesive Dressings-2 each 9. End Cap 10. suturewith curved needle, 11. scalpel blade USFDA DCGI marked							
20	Double Lumen split tip tunneled hemodialysis catheter kit pediatric size 1. Tip to cuff length 8, 10, 12, 15, 19, 23 cm 2. 7.5Fr-10 Fr Catheter 3. 8F-10Fr AirGuard Valved Introducer with Peel Away Sheath/Dilator 4. Dilator 5. Tunneler 6. 2 End Caps 7. J-Tip Guidewire 0.032" 8. 18 Gauge Introducer Needle 9. 2 Adhesive Dressing USFDA DCGI marked							
21	Disposable Core tissue Biopsy needle compatible with Magnum Biopsy instrument size: 16G X 16Cm USFDA DCGI marked							
22	Disposable Core tissue Biopsy needle compatible with Magnum Biopsy instrument size: 18G X 16Cm USFDA DCGI marked							
23	Disposable Core tissue Biopsy needle compatible with Magnum Biopsy instrument size: 18G X 10Cm USFDA/ EUROPEAN CE approved and DCGI marked							
24	Sterilit oil spray to reduce friction and corrosion due to friction for instruments (300ml)							
25	Bacterial Filter for Macquet Ventilator							
26	RBC Leukocyte reduction filters (lab side with transfer bag) 3rd generation							
27	Micro-puncture kit							
28	PTA & INJECTION COMBINED · 1 Should have 3 port system. · 2 Should be available in sizes, Dia(5mm,6mm,7mm) with nominal pressure of 14atm upto burst pressure of 25atm. · 3 Should enable simultaneous angioplasty and variety of injections at any time. · 4 Should combine the functionality of high-pressure balloon catheter and diagnostic catheter · 5 Should be able to perform an entire procedure both angioplasty and injection of diagnostic or therapeutic fluids. 6 Should be able to perform reflex angiography. 7 The balloon should have supervision Technology USFDA/ EUROPEAN CE approved and DCGI marked							
29	Tenckhoff peritoneal dialysis catheter, 2cuffs , 42cm, Straight with introducer set for percutaneous insertion USFDA/ EUROPEAN CE approved and DCGI marked							
30	Tenckhoff peritoneal dialysis catheter, 2cuffs , 31cm, Straight with introducer set for percutaneous insertion USFDA/ EUROPEAN CE approved and DCGI marked							
31	Tenckhoff peritoneal dialysis catheter, 2cuffs , 37cm, Straight with introducer set for percutaneous insertion USFDA/ EUROPEAN CE approved and DCGI marked							
32	Tenckhoff peritoneal dialysis catheter, 2cuffs , 46cm, Straight with introducer set for percutaneous insertion USFDA/ EUROPEAN CE approved and DCGI marked							
33	Tenckhoff peritoneal dialysis catheter, single cuffs , 31cm, Straight with introducer set for percutaneous insertion USFDA/ EUROPEAN CE approved and DCGI marked							
34	Tenckhoff peritoneal dialysis catheter, single cuffs , 37cm, Straight with introducer set for percutaneous insertion USFDA/ EUROPEAN CE approved and DCGI marked							
35	Tenckhoff peritoneal dialysis catheter, single cuffs , 42cm, Straight with introducer set for percutaneous insertion USFDA/ EUROPEAN CE approved and DCGI marked							
36	Acute semirigid peritoneal dialysis catheter with Y transfer set Adult size USFDA/ EUROPEAN CE approved and DCGI marked							

S.No	Description of consumable	Pack Size	Unit	Pack size offered	Name of Brand/Company Quoted	Indian/Imported/Country of Origin	Catalogues Yes/No	Sample Yes/No
37	Acute semirigid peritoneal dialysis catheter with Y transfer set pediatric size USFDA/ EUROPEAN CE approved and DCGI marked							
38	Pediatric acute peritoneal dialysis set Including 1. Polyurethane catheter-8.5 fr, length 8 cm 2. Guide wire -0.35 inch length -30 cm 3. introducer needle 18T gage 4. connecting tube-30 cm 5. dilator 6. retention disc with the pull silicone USFDA/ EUROPEAN CE approved and DCGI marked							
39	Locking titanium adapter for peritoneal dialysis catheter USFDA/ EUROPEAN CE approved and DCGI marked							
40	Minicap extended life PD transfer set with twist clamp USFDA/ EUROPEAN CE approved and DCGI marked							
41	Peritoneal dialysis fluids 1.5 % 2000ml bag Each 100 ml – dextrose anhydrous 1.5gm, sodium lactate 448 mg, Na Cl 538 mg, Cacl 25.7 mg, Mgcl- 5.08 mg with integrated asymmetric Y set with or without minicap with colour code							
42	Peritoneal dialysis fluids 2.5 % 2000ml bag Each 100 ml – dextrose anhydrous 2.5gm, sodium lactate 448 mg, Na Cl 538 mg, Cacl 25.7 mg, Mgcl- 5.08 mg with integrated asymmetric Y set with or without minicap with colour code							
43	Peritoneal dialysis fluids 7.5 % 2000ml bag Each 100 ml – icodextrin 7.5gm, sodium lactate 448 mg, Na Cl 538 mg, Cacl 25.7 mg, Mgcl- 5.08 mg with water for injection Q S							
44	Peritoneal dialysis fluids 1.5 % 5000ml bag Each 100 ml – dextrose anhydrous 1.5gm, sodium lactate 448 mg, Na Cl 538 mg, Cacl 25.7 mg, Mgcl- 5.08 mg with							
45	Peritoneal dialysis fluids 2.5 % 5000ml bag Each 100 ml – dextrose anhydrous 2.5gm, sodium lactate 448 mg, Na Cl 538 mg, Cacl 25.7 mg, Mgcl- 5.08 mg with water for injection Q S							
46	16 Fr Tearway introducer sheath for Peritoneal Dialysis							
47	Multifiltrate cassette compatible with Fresenius CRRT machine USFDA/ EUROPEAN CE approved and DCGI marked							
48	Substitution line compatible with Fresenius CRRT machine USFDA/ EUROPEAN CE approved and DCGI marked							
49	Dialysate line compatible with Fresenius CRRT machine USFDA/ EUROPEAN CE approved and DCGI marked							
50	Ultra-flux filter compatible with Fresenius CRRT machine USFDA/ EUROPEAN CE approved and DCGI marked							
51	Filtrate bag compatible with Fresenius CRRT machine USFDA/ EUROPEAN CE approved and DCGI marked							
52	Female spike adaptor 4 each compatible with Fresenius CRRT machine USFDA/ EUROPEAN CE approved and DCGI marked							
53	Solution for continuous hemofiltration and hemodialysis, potassium free, Bicarbonate 30-35 meq/L, Calcium 2.5 - 3.5 meq/L with Dextrose, 5000ml/bag USFDA/ EUROPEAN CE approved and DCGI marked							
54	Solution for continuous hemofiltration and hemodialysis, potassium free, Bicarbonate 30-35 meq/L, Calcium 2.5 - 3.5 meq/L without Dextrose, 5000ml/bag USFDA/ EUROPEAN CE approved and DCGI marked							
55	Solution for continuous hemofiltration and hemodialysis, potassium free, Bicarbonate 30-35 meq/L, with Dextrose, without calcium, 5000ml/bag USFDA/ EUROPEAN CE approved and DCGI marked							